



(2016) GRJR Zi (099)



China Accreditation
International mutual
accreditation inspection
TESTING
CNAS L1499

Inspection Report

Inspection Report No.: No. 2020 (dust) - 866

Sample Name: Protective Mask

Entrusted by: Qingdao Huaren Medical Product Co., Ltd.

Inspection Type: Entrusting

National Quality Supervision and Inspection Center for Labor Protection Articles (Beijing)

Seal: National Quality Supervision and Inspection Center for Labor Protection Articles
(Beijing)

**National Quality Supervision and Inspection Center for
Labor Protection Articles (Beijing)**



Inspection Report No.: No. 2020 (dust) – 866

Page 1 of 2


Sample name	Protective mask	Trademark	Ren
Specification	folded type	Protection Level	-----
Name of entrusting unit	Qingdao Huaren Medical Product Co., Ltd.	Tel. of entrusting unit	0532-58070299
Address of entrusting unit	No. 187, Zhuzhou Road, Laoshan District, Qingdao, Shandong Province		
Name of manufacturer	Qingdao Huaren Medical Product Co., Ltd.	Tel. of manufacturer	0532-58070299
Address of manufacturer	No. 187, Zhuzhou Road, Laoshan District, Qingdao, Shandong Province		
Sample quantity	10 pcs	Production date	02/03/2020
Sample status	Complete	Sample arrival date	04/03/2020
Sample features	Whitefolded type mask	Sample sender	Wang Huan
Inspection basis	GB2626-2006 Respiratory protective equipment – Non-powered air-purifying particle respirator		
Inspection item	Filtration efficiency, inhalation resistance, respiratory resistance		
Inspection conclusion	<p>In terms of the sample, the inspected item data are verified to comply with (KN95) standard requirements in accordance with GB 2626-2006 <i>Respiratory protective equipment – Non-powered air-purifying particle respirator</i>.</p> <p>(Special Seal of Inspection Report) Seal: Special Seal of Inspection Report, National Quality Supervision and Inspection Center for Labor Protection Articles (Beijing) Issuing date: 10/03/2020</p>		
Remarks	The sample is not preprocessed.		
Approved by:		Verified by:	
		Main inspector:	

**National Quality Supervision and Inspection Center for Labor Protection
Articles (Beijing)**

Inspection Report (Summary of Inspection Results)

Inspection Report No.: No. 2020 (dust) – 866

Page 2 of 2

S.N.	Inspection item	Standard Article No.	Standard Requirement	Measured Result			Current Conclusion	Remarks	
1	Filtration efficiency	5.3	KN90 ≥ 90.0% KN95 ≥ 95.0% KN100 ≥ 99.97% Test of air flow: Single filter element: (85±4) L/min	Non preprocessed		Initial filtration efficiency (η)	Loading filtration efficiency (η)	Meet requirements in the standard (KN95)	--
					1#	99.2%	99.0%		
					2#	99.2%	98.9%		
					3#	99.2%	98.9%		
					4#	99.2%	98.9%		
					5#	99.2%	98.9%		
					6#	99.2%	98.9%		
					7#	99.2%	99.0%		
					8#	99.2%	98.9%		
					9#	99.2%	98.9%		
			KN type: Environment temperature: (25±5)°C Relative humidity: (30±10)%	Environment temperature: 24°C Relative humidity: 32%					
2	Inhalation resistance	5.5	Total inhalation resistance of each sample shall not exceed 350Pa.	Non preprocessed	3#	88Pa		Meet requirements in the standard	--
					4#	94Pa			
3	Respiratory resistance	5.5	Total respiratory resistance of each sample shall not exceed 250Pa.	Non preprocessed	3#	94Pa		Meet requirements in the standard	--
					4#	94Pa			
Sample picture									
Remarks			-----						

Inspection date: 04/03/2020 ~ 05/03/2020

----- END -----

Seal: Special Seal of Inspection Report, National Quality Supervision and Inspection Center for Labor Protection Articles (Beijing)

Cautions

1. This report shall become invalid without special seal of inspection report or official seal of the inspection unit.
2. The copy shall be invalid without special seal of inspection report or official seal of the inspection unit.
3. The report is invalid without signature or seal of main inspector, verifier or approver.
4. The report is invalid in case of any alteration.
5. The entrusted inspection shall only be subject to the sample.
6. Any part of this report is prohibited to copy without written consents of the laboratory.

Seal: Special Seal of Inspection Report, National Quality Supervision and Inspection Center for Labor Protection Articles (Beijing)

Account name: Beijing Municipal Institute of Labour Protection

Opening bank: ICBC Beijing Taoranting Subbranch

A/C No.: 0200049109024901958

Address: No. 55, Taoranting Road, Xicheng District, Beijing

P.C.: 100054

Tel.: 010-63520770

Fax: 010-63520770

Email: bjlb63520770@vip.163.com



160010260248



(2016)国认监认字(099)号



中国认可
国际互认
检测
TESTING
CNAS L1499

检 验 报 告

检验报告编号：No. 2020(尘) - 866



样品名称： 防护口罩


委托单位： 青岛华仁医疗用品有限公司

检验类别： 委 托

国家劳动保护用品质量监督检验中心(北京)



国家劳动保护用品质量监督检验中心（北京）
检 验 报 告



检验报告编号: No. 2020(尘) - 866


共 2 页 第 1 页

样品名称	防护口罩	商 标	仁
规格型号	折叠式	防护级别	—————
委托单位名称	青岛华仁医疗用品有限公司	委托单位电话	0532-58070299
委托单位地址	山东省青岛市崂山区株洲路 187 号		
生产单位名称	青岛华仁医疗用品有限公司	生产单位电话	0532-58070299
生产单位地址	山东省青岛市崂山区株洲路 187 号		
样品数量	10 个	生产日期	2020.03.02
样品状态	完好	到样日期	2020.03.04
样品特征	白色折叠口罩	送 样 者	王欢
检验依据	GB 2626-2006 《呼吸防护用品 自吸过滤式防颗粒物呼吸器》		
检验项目	过滤效率、吸气阻力、呼气阻力		
检 验 结 论	<p>该样品经检验,依据 GB 2626-2006 《呼吸防护用品 自吸过滤式防颗粒物呼吸器》,判定所检项目数据符合 (KN95) 标准要求。</p> <p style="text-align: center;">(检验报告专用章)</p> <p style="text-align: center;">签发日期: 2020年3月10日</p>		
备 注	样品未做预处理		
批 准:	陈辉为	审 核:	尹超
		主 检:	王欢

国家劳动保护用品质量监督检验中心（北京）
检验报告（检验结果汇总）

检验报告编号：No. 2020(尘) - 866

共 2 页 第 2 页

序号	检验项目	标准条款号	标准要求	实测结果			本项结论	备注	
					初始过滤效率 (η)	加载过滤效率 (η)			
1	过滤效率	5.3	KN90 \geq 90.0% KN95 \geq 95.0% KN100 \geq 99.97% 测试气流量： 单过滤元件： (85 \pm 4) L/min	未做 预处理	1#	99.2 %	99.0 %	符合 标准要求 (KN95)	---
					2#	99.2 %	98.9 %		
					3#	99.2 %	98.9 %		
					4#	99.2 %	98.9 %		
					5#	99.2 %	98.9 %		
					6#	99.2 %	98.9 %		
					7#	99.2 %	99.0 %		
					8#	99.2 %	98.9 %		
					9#	99.2 %	98.9 %		
					10#	99.2 %	99.0 %		
KN 类： 环境温度：(25 \pm 5) $^{\circ}$ C 相对湿度：(30 \pm 10) %				环境温度： 24 $^{\circ}$ C 相对湿度： 32 %					
2	吸气阻力	5.5	每个样品的总吸气阻力应 \leq 350Pa。	未做 预处理	3#	88 Pa		符合 标准要求	---
					4#	94 Pa			
3	呼气阻力	5.5	每个样品的总呼气阻力应 \leq 250Pa。	未做 预处理	3#	94 Pa		符合 标准要求	---
					4#	94 Pa			
样品照片									
备注		<div style="text-align: right;">  </div>							

检验日期：2020 年 03 月 04 日至 2020 年 03 月 05 日

----- 结 束 -----

注 意 事 项

1. 报告无检验报告专用章或检验单位公章无效。
2. 复制报告未加盖检验报告专用章或检验单位公章无效。
3. 报告无主检、审核、批准人签章无效。
4. 报告涂改无效。
5. 委托检验仅对来样负责。
6. 未经实验室书面批准，不得部分复印本报告。



开 户 名：北京市劳动保护科学研究所

开 户 行：工行北京陶然亭支行

帐 号：0200049109024901958

地 址：北京市西城区陶然亭路 55 号

邮 编：100054

电 话：010—63520770

传 真：010—63520770

邮 箱：bjlb63520770@VIP.163.com